




Cell-VU[®] Prestained Morphology Slides

DRM-900

Direction for use

1. Apply a small drop (approximately ~2-4 microliters) of specimen on the center of the slide.
2. Place a standard coverglass (24 x 24 mm or 24 x 40 mm) on top of the slide, exerting slight pressure on the coverglass.
Note: Do not slide the coverglass as this will remove the stain.
Note: You can place covered slide coverglass side down on paper towel and press gently.
3. Allow slide to stand at room temperature for 15 minutes before microscopic observation is made. *(Place loaded slide on heating plate for 1 minute if you do not want to wait.)*



Expected Findings:

These slides are for evaluating sperm head/acrosome size, differentiation of white blood cells in peripheral blood and cytology of body fluids.

Warnings:

EDTA is the recommended anticoagulant for venous blood. Oxalae, flouride and heparinized blood are not suitable.

Notes:

For best results read the same day.

1997: Millennium Sciences, Inc.
Cell-VU is a registered trademark of Millennium Sciences, Inc.